

**Miller, Diane M. (CDC/NIOSH/EID)**

---

**From:**  
**Sent:** Wednesday, November 26, 2008 12:16 PM  
**To:** NIOSH Docket Office (CDC)  
**Cc:** Chen, Jihong (Jane) (CDC/NIOSH/EID) (CTR); Doyle, Glenn (CDC/NIOSH/EID)  
**Subject:** 034-A - SCBA-EOSTI Comments

**Name**

**Organization**

**Email**

**Address**

**Comments**

I believe the only modificatios to the EOSTI should be an audible warning in addition to the now existing Vibrating face piece regultors. As an instructor I have found that students in training environments do not recognize when another members face piece is vibrating and need to have additional warnings to alert them of the possiility of a problem with their partner or other members working on the fireground.